

MULTICENTRIC STUDY: IMPROVEMENT IN URINARY INCONTINENCE AND QUALITY OF LIFE

SAFETY AND EFFICACY OF A NON-INVASIVE HIGH-INTENSITY FOCUSED ELECTROMAGNETIC FIELD (HIFEM®) DEVICE FOR TREATMENT OF URINARY INCONTINENCE AND ENHANCEMENT OF QUALITY OF LIFE

Julene B. Samuels, MD,¹ Andrea Pezzella, MD,² Joseph Berenholz, MD,³
and Red Alinsod, MD⁴

1. Julene B. Samuels, MD, FACS, Louisville, KY

2. Southern Urogynecology: Center for Female Pelvic Medicine and Reconstructive Surgery, West Columbia, SC

3. The Laser Vaginal Rejuvenation Institute Of Michigan Farmington Hills, MI

4. South Coast Urogynecology, Laguna Beach, CA

Published in *Lasers in Surgery and Medicine*, 2019

HIGHLIGHTS

- Patients reached on average **64.42% improvement in ICIQ-UI score** and **53.68% decrease in absorbent pad usage**.
- **21 (34.43%) subjects were cured** (zero score in ICIQ-UI and free from UI symptoms at follow-up).
- **64.29% fewer patients experienced leakage while sleeping**.
- **19 patients (44.19%) reported they were not using pads anymore**.

Parameter	ICIQ-SF	Absorbent Pads
Number of evaluated subjects	61	43
Baseline	10.57±4.22	2.47±2.25
After 6th Treatment	5.33±3.97	1.35±1.74
<i>Difference Before & Follow-up</i>	5.25±4.02	1.12±1.80
<i>Average Improvement</i>	49.93%	43.80%
3-month Follow-Up	4.16±4.04	1.19±1.91
<i>Difference Before & Follow-up</i>	6.41±3.75	1.28±1.83
<i>Average Improvement</i>	64.42%	53.68%